



COMMONWEALTH of VIRGINIA

DEPARTMENT OF FORENSIC SCIENCE

CENTRAL LABORATORY
A Nationally Accredited Laboratory

700 NORTH 5TH STREET
RICHMOND, VIRGINIA 23219
(804) 786-6971

CERTIFICATE OF INSTRUMENT ACCURACY

INSTRUMENT: Intox EC/IR II
SERIAL NUMBER: 010446
CERTIFICATION DATE: November 22, 2013
DATE RECERTIFICATION DUE: May 21, 2014

The above listed instrument was found to meet all requirements for accuracy and performance established by the Department of Forensic Science. This certification is valid from the above listed certification date until the above listed date for recertification.


Melissa Kennedy

Program Manager, Calibration & Training

CITY OF RICHMOND
COMMONWEALTH OF VIRGINIA

The foregoing was acknowledged before me this 2nd day of
December, 2013 by Melissa Kennedy

Ciera N. Boxley, Notary Public



My commission expires: August 31, 2014

504

Department of Forensic Science
Breath Alcohol Section
Instrument Maintenance History
1/1/1990 To 11/25/2013

HMS
NW

Instrument Serial Number: 010446	As of 25-Nov-13
---	------------------------

Maintenance Date	Certification Date	Technician Initials	Remarks
22-Nov-13	22-Nov-13	HMS (18910)	REPLACED PRINTER RIBBON, FIRMWARE/SOFTWARE UPGRADE, CALIBRATED AND CERTIFIED, REPLACED DRY GAS STANDARD.
31-May-13	31-May-13	HMS (18910)	CERTIFIED
04-Apr-13	18-Dec-12	TLN (25387)	REPLACED UPS BACK-UP BATTERY.
18-Dec-12	18-Dec-12	HMS (18910)	CERTIFIED
15-Nov-12	15-Nov-12	HMS (18910)	CERTIFIED
09-Nov-12	09-Nov-12	TLN (25387)	FIRMWARE/SOFTWARE UPGRADE. CALIBRATED AND CERTIFIED.
09-Dec-11	09-Dec-11	DBH (21250)	CERTIFIED
21-Oct-11	21-Oct-11	DBH (21250)	REPLACED DRY GAS STANDARD, CERTIFIED
25-Jul-11	25-Jul-11	DBH (21250)	CERTIFIED
15-Feb-11	15-Feb-11	RJP (22428)	CERTIFIED.
27-Jul-10	03-Jun-10	RJP (22428)	REPLACED DRY GAS STANDARD.
03-Jun-10	03-Jun-10	RJP (22428)	REPLACED DRY GAS STANDARD. CERTIFIED
12-Apr-10	17-Dec-09	RJP (22428)	REPLACED DRY GAS STANDARD.
04-Feb-10	17-Dec-09	RJP (22428)	REPLACED DRY GAS STANDARD.
18-Dec-09	17-Dec-09	SED (24178)	CORRECTED DRY GAS STANDARD TANK NUMBER IN INSTRUMENT MEMORY.
17-Dec-09	17-Dec-09	RJP (22428)	REPLACED DRY GAS STANDARD. CERTIFIED
09-Oct-09	09-Oct-09	RJP (22428)	FIRMWARE/SOFTWARE UPGRADE. CERTIFIED.
26-Jun-09	26-Jun-09	HMS (18910)	FIRMWARE/SOFTWARE UPGRADE, CERTIFIED.
14-May-09	14-May-09	RJP (22428)	REPLACED TANK TOP LID. CERTIFIED.
13-Nov-08	13-Nov-08	RJP (22428)	CERTIFIED.
21-Oct-08	21-Oct-08	NMJ (22392)	CALIBRATED, CERTIFIED AND ACCEPTED.

INTOX EC/IR II
Quality Assurance Worksheet

Handwritten initials and signature in the top right corner.

Instrument Serial Number 010446 Worksheet Start Date 11/22/2013

Location Meherrin River Regional Jail

Address 9000 Boydton Plank Rd, Alberta, VA 23821

DFS Technician Heather Stanton License No. 18910

☐ Laboratory ☒ On-Site

Site Specifications

Yes No

1 Secured Environment:

Location must not be accessible to the general public.

☒ ☐

2 Environment free from contaminants

☒ ☐

3 Dedicated telephone line:

The phone line should be used solely by the instrument.

☒ ☐

4 Adequate Ventilation:

Enough space must be provided for the instrument and the printer to allow for adequate ventilation.

☒ ☐

5 Stable temperature controls:

The instrument should not be exposed to direct sources of temperature variation.

☒ ☐

6 115VAC power source:

Whenever possible instrument power circuit should not be shared with refrigerators, copy machines or other high amperage devices.

☒ ☐

7 Room Condition:

The site should be free of excessive moisture, heat and dust.

☒ ☐

	Old serial no.	New serial no.	Supplies	
E-Seek			Mouthpieces	
Keyboard			Certificates of Analysis	
Printer			Operator Worksheet	
UPS			Other:	

50
2 pads

Maintenance (if any)
Replaced printer ribbon. Firmware/software upgrade.

Comments (if any)

Instrument Serial Number 010446

Certification Date 11/22/2013

Instrument Barometer (mm HG) 756

Reference Barometer (mm HG) 758

Reference Barometer(RB)Serial # 009113

RB Calibration Due 5/8/2014

Standard (sea level)	PA Target	minimum	maximum	Sample 1	
0.250	0.248	0.240	0.255	0.246	
Precision		sample min	sample max	Sample 2	0.246
0.001		0.245	0.246	Sample 3	0.245

Standard (sea level)	PA Target	minimum	maximum	Sample 1	
0.150	0.149	0.144	0.153	0.148	
Precision		sample min	sample max	Sample 2	0.148
0		0.148	0.148	Sample 3	0.148

Standard (sea level)	PA Target	minimum	maximum	Sample 1	
0.080	0.080	0.077	0.083	0.079	
Precision		sample min	sample max	Sample 2	0.079
0		0.079	0.079	Sample 3	0.079

Standard (sea level)	PA Target	minimum	maximum	Sample 1	
0.020	0.020	0.017	0.023	0.019	
Precision		sample min	sample max	Sample 2	0.019
0		0.019	0.019	Sample 3	0.019

all measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

☒ Calibrated☒ Certified☒ Replaced dry gas standard (+O-ring)

Dry gas standard Lot No. (with tank no.)

AG320502-06

☒ Measurement Assurance Check☒ Run a Breath Test (Instrument Test)

(successful completion denotes satisfactory condition of the item)

Standard (sea level)	PA Target	minimum	maximum	Sample 1	
0.100	0.099	0.096	0.102	0.099	
Precision		sample min	sample max	Sample 2	0.099
0		0.099	0.099	Sample 3	0.099

DFS Technician

Heath W. Stark

Date

11/25/13

Issuing Analyst

Melissa Kennedy

Date

12/1/13

Intox EC/IR II QA Worksheet
Issued by Calibration and Training Program Manager
Issue Date: 4-March-2013

DFS Document 250-F113
Revision 6
Page 2 of 2




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 11/22/2013
TEST EQUIPMENT NUMBER 010446		

RESULTS: TIME SAMPLE TAKEN 12:34 EST

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 11/22/2013
TEST EQUIPMENT NUMBER 010446		

RESULTS: TIME SAMPLE TAKEN 12:34 EST

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

*Replaced printer ribbon.
HMS
11/22/13*

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report
Report Generated 22 Nov 2013 at 13:33

TZ
HPS
NHL

Test Results

Instrument Serial Number 010446

Test # 001903 Subject Test

Test Location 1 Department of
Test Date 22 Nov 2013

Test Location 2 Forensic Science
Test Time 12:28
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M
License Number 18910

Card Serial Number 118910

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 22 Nov 2013 End Time 12:35

Result Time 12:34

Result Date 22 Nov 2013 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 12:28

Data Type BLK

Sample Value 0.000

Sample Time 12:29

Data Type CHK

Sample Value 0.099

Sample Time 12:29

Data Type BLK

Sample Value 0.000

Sample Time 12:31

Data Type SUBJ

Sample Value 0.000

Sample Time 12:31

Data Type BLK

Sample Value 0.000

Sample Time 12:32

Data Type SUBJ

Sample Value 0.000

Sample Time 12:34

Data Type BLK

Sample Value 0.000

Sample Time 12:35

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG217702-16

Standard Expiration Date 06/25/2014

Tank Pressure 772

Barometric Pressure 756 mmHg

Blow Sample Number 1 Blow Duration 3.14 sec

Blow Volume 1739 cc End-of-Blow Time 12:31

Blow Sample Number 2 Blow Duration 3.29 sec

Blow Volume 1725 cc End-of-Blow Time 12:34

Tamper Evident Stamp 777b7474

Test Status Code 0

Test Status Success

HMS

MSL

Intox EC/IR-II: Certification

Department of Forensic Science

Serial Number: 010446 Test Number: 1904
Test Date: 11/22/2013 Test Time: 12:38 EST

Technician Initials: HMS

Dry Gas Target: 0.100

Lot Number: AG217702-16 Exp Date: 06/25/2014

Tank Pressure: 767 psi Barometric Pressure: 756 mmHg
System Check: Passed

Test	g/210L	Time
DIAG	Pass	12:39
BLK	0.000	12:40
CHK	0.099	12:40
BLK	0.000	12:42
CHK	0.099	12:42
BLK	0.000	12:44
CHK	0.099	12:44
BLK	0.000	12:46

Success

Standard Value (at sea level): 0.100 g/210L

Certification Status: Success

One point certification done prior to firmware/software upgrade due to certification warning on display.

HMS
11/22/13

hms
wax

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010446 Test Number: 1905

Test Date: 11/22/2013 Test Time: 12:47 EST

Dry Gas Target: 0.100

Lot Number: AG217702-16 Exp Date: 06/25/2014

Tank Pressure: 757 psi Barometric Pressure: 756 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:48
CHK	0.099	12:49
BLK	0.000	12:50
CHK	0.098	12:51
BLK	0.000	12:52
CHK	0.099	12:53

Test Status: *Success*

Calibration CRC: 945A3DD9

Handwritten signature

Handwritten signature

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010446 Test Number: 1906

Test Date: 11/22/2013 Test Time: 12:56 EST

Dry Gas Target: 0.298

Lot Number: AG322702-01 Exp Date: 08/15/2015

Tank Pressure: 134 psi Barometric Pressure: 756 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:57
CHK	0.298	12:58
BLK	0.000	12:59
CHK	0.298	13:00
BLK	0.000	13:02
CHK	0.297	13:02

Test Status: *Success*

Calibration CRC: 945A3DD9

Handwritten initials/signature in the top right corner.

Intox EC/IR-II: Software Version

Department of Forensic Science

Serial Number: 010446

System Date: 11/22/2013

System Time: 13:24 EST

Software Version: VS00333-K

Software CRC: 56745BCA

Firmware Version: VS00171-CL

Firmware CRC: 7107756B

Has
MSL

Intox EC/IR-II: Calibration

Department of Forensic Science

Serial Number: 010446 Test Number: 1907

Test Date: 11/22/2013 Test Time: 13:25 EST

Dry Gas Target: 0.099

Lot Number: 0651210013-06 Exp Date: 04/01/2014

Tank Pressure: 580 psi Barometric Pressure: 755 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	13:26
CAL	0.099	13:27

Test Status: *Success*

Calibration CRC: 8312375D

HMS
AWK

Intox EC/IR-II: Certification

Department of Forensic Science

Serial Number: 010446 Test Number: 1908
Test Date: 11/22/2013 Test Time: 13:29 EST

Technician Initials: HMS

Dry Gas Target: 0.248

Lot Number: 2961225012-08 Exp Date: 11/01/2014
Tank Pressure: 242 psi Barometric Pressure: 755 mmHg
System Check: Passed

Test	g/210L	Time
DIAG	Pass	13:30
BLK	0.000	13:31
CHK	0.246	13:31
BLK	0.000	13:33
CHK	0.246	13:34
BLK	0.000	13:35
CHK	0.245	13:36
BLK	0.000	13:38

Success

Standard Value (at sea level): 0.250 g/210L

Certification Status: Continued

AND

WXX

Intox EC/IR-II: Certification

Department of Forensic Science

Serial Number: 010446 Test Number: 1909
Test Date: 11/22/2013 Test Time: 13:39 EST

Technician Initials: HMS

Dry Gas Target: 0.149

Lot Number: 2951215012-02 Exp Date: 11/01/2014
Tank Pressure: 718 psi Barometric Pressure: 755 mmHg
System Check: Passed

Test	g/210L	Time
DIAG	Pass	13:39
BLK	0.000	13:40
CHK	0.148	13:41
BLK	0.000	13:42
CHK	0.148	13:43
BLK	0.000	13:44
CHK	0.148	13:45
BLK	0.000	13:46

Success

Standard Value (at sea level): 0.150 g/210L

Certification Status: Continued

Handwritten initials: HMS and RSK

Intox EC/IR-II: Certification

Department of Forensic Science

Serial Number: 010446 Test Number: 1910
Test Date: 11/22/2013 Test Time: 13:48 EST

Technician Initials: HMS

Dry Gas Target: 0.080

Lot Number: 0771308012-04 Exp Date: 05/01/2015
Tank Pressure: 786 psi Barometric Pressure: 755 mmHg
System Check: Passed

Test	g/210L	Time
DIAG	Pass	13:48
BLK	0.000	13:49
CHK	0.079	13:50
BLK	0.000	13:51
CHK	0.079	13:52
BLK	0.000	13:53
CHK	0.079	13:54
BLK	0.000	13:55

Success

Standard Value (at sea level): 0.080 g/210L

Certification Status: Continued

HMS
PDL

Intox EC/IR-II: Certification

Department of Forensic Science

Serial Number: 010446 Test Number: 1911
Test Date: 11/22/2013 Test Time: 13:56 EST

Technician Initials: HMS

Dry Gas Target: 0.020

Lot Number: 0771302011-04 Exp Date: 05/01/2015
Tank Pressure: 884 psi Barometric Pressure: 755 mmHg
System Check: Passed

Test	g/210L	Time
DIAG	Pass	13:57
BLK	0.000	13:57
CHK	0.019	13:58
BLK	0.000	13:59
CHK	0.019	14:00
BLK	0.000	14:01
CHK	0.019	14:02
BLK	0.000	14:03

Success

Standard Value (at sea level): 0.020 g/210L

Certification Status: Success

AMS
Nell

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010446 Test Number: 1912

Test Date: 11/22/2013 Test Time: 14:05 EST

Dry Gas Target: 0.298

Lot Number: AG322702-01 Exp Date: 08/15/2015

Tank Pressure: 129 psi Barometric Pressure: 755 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	14:06
CHK	0.298	14:07
BLK	0.000	14:08
CHK	0.298	14:09
BLK	0.000	14:11
CHK	0.299	14:12

Test Status: *Success*

Calibration CRC: 8312375D

Handwritten marks:
HMS
NDE

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010446 Test Number: 1913

Test Date: 11/22/2013 Test Time: 14:15 EST

Dry Gas Target: 0.099

Lot Number: AG320502-06 Exp Date: 07/24/2015

Tank Pressure: 1120 psi Barometric Pressure: 755 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	14:15
CHK	0.099	14:16
BLK	0.000	14:18
CHK	0.099	14:19
BLK	0.000	14:20
CHK	0.099	14:21

Test Status: *Success*

Calibration CRC: 8312375D




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

AMS
msc

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 11/22/2013
TEST EQUIPMENT NUMBER 010446		

RESULTS: TIME SAMPLE TAKEN 14:28 EST
SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 25 Nov 2013 at 09:32

Handwritten initials/signature

Test Results

Instrument Serial Number 010446

Test # 001914 Subject Test

Test Location 1 Department of

Test Date 22 Nov 2013

Test Location 2 Forensic Science

Test Time 14:22

Remote/Local Local

Test Location 3

System Check Passed

Operator's Last Name STANTON

Agency DFS Central Lab

Card Serial Number 118910

Subject's Last Name INSTRUMENT

Subject's Middle Initial

Driver's License Number

Driver's License State

Operator's First Name HEATHER

License Number 18910

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's First Name TEST

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Expiration 00/00/0000

Court Name DFS

End Date 22 Nov 2013

End Time 14:30

Result Time 14:28

Result Date 22 Nov 2013

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 14:22

Data Type BLK

Sample Value 0.000

Sample Time 14:23

Data Type CHK

Sample Value 0.099

Sample Time 14:24

Data Type BLK

Sample Value 0.000

Sample Time 14:25

Data Type SUBJ

Sample Value 0.000

Sample Time 14:26

Data Type BLK

Sample Value 0.000

Sample Time 14:27

Data Type SUBJ

Sample Value 0.000

Sample Time 14:28

Data Type BLK

Sample Value 0.000

Sample Time 14:29

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG320502-06

Standard Expiration Date 07/24/2015

Tank Pressure 1120

Barometric Pressure 755 mmHg

Blow Sample Number 1

Blow Duration 3.33 sec

Blow Volume 1751 cc

End-of-Blow Time 14:26

Blow Sample Number 2

Blow Duration 3.32 sec

Blow Volume 1713 cc

End-of-Blow Time 14:28

Tamper Evident Stamp 7abc4f9d

Test Status Code 0

Test Status Success